

Supplementary Material

Table S1. Raw data for limit of detection (LOD) determination by probit analysis

Concentration (copies/mL)	Positive/Total	Detection Rate (%)
2500	16/16	100
1000	16/16	100
500	15/16	93.75
250	12/16	75
100	8/16	50